



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2024/B/4929872
Dated/दिनांक : 11-05-2024

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	01-06-2024 11:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	01-06-2024 11:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	120 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Health And Family Welfare
Department Name/विभाग का नाम	Department Of Health And Family Welfare
Organisation Name/संगठन का नाम	All India Institute Of Medical Sciences (aiims)
Office Name/कार्यालय का नाम	Jammu
Total Quantity/कुल मात्रा	3
Item Category/मद केटेगरी	Advanced Dental Chair (Q3)
Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)	40 Lakh (s)
OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)	100 Lakh (s)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	3 Year (s)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance/विगत प्रदर्शन	80 %

Bid Details/बिड विवरण	
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	252000

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	62

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

ED & CEO
All India Institute of Medical Sciences (AIIMS), VIJAYPUR, JAMMU
(Shatki Kumar Gupta)

Splitting/विभाजन

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता
--

Yes

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता

Yes

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.
[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.
5. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY.The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023
[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.
6. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 80% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid

in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Pre Bid Detail(s)

Pre-Bid Date and Time	Pre-Bid Venue
17-05-2024 14:00:00	Dentistry OPD, First Floor, Tower6, AllMS, Vijaypur, Jammu 184120

Advanced Dental Chair (3 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Performance Parameters	Chair along with complete unit is suitable for	Right handed operator, Left handed operator, It is a convertible chair/unit for Left and Right handed operators
	Chair has a baseless design	Yes, No
	Chair has provision for swivel of Dentist Arm Left/Right, with more than 300 degrees movement	Yes, No
	Chair has safety brake system at base and backrest while going down, for patient's exit position	Yes, No
	Lifting capacity of chair	Less than 150 Kg, 150 - 180 Kg, 180 - 200 Kg, More than 200 Kg
	Operation	Electrical, Mechanical
	Upholstery is Seamless and wipe disinfectable	Yes, No
	No of programs available	4, 5, 6, more than 6
	Program for last working position	Available, Not Available
	Program for return to 'zero position'	Available, Not Available
	Synchronized backrest and footrest moment	Yes, No
	Headrest with multi-axis rotation and height adjustable	No, Yes(mechanical), Yes (Pneumatic)
	Headrest with pneumatically controlled stop	Yes, No

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	LED operating light	Yes (with touch free operation), Yes (with manual operation)
	LED operating light intensity	25,000-30,000 lux, 30,000-40,000 lux, 40,000-50,000 lux, more than 50,000 lux, NA (if no led light supplied)
	LED operating light is colour corrected	Yes, No
	Rinsing water/Cup water heating facility available	Yes, No
	Spittoon	High finished Porcelain, Toughened glass
	Spittoon comes with motorized motion and programmed cup filler and flush function	Yes, No
	Spittoon bowl removable	Yes, No
	Spittoon bowl swiveable	Yes, No
	Spittoon outlet has facilities of Biomedical waste treatment /water treatment	Yes, No
	Coupling for light cure unit available on assistant side also	Yes, No
	Three way syringe available on assistant side also	Yes, No
	High and low suction available on assistant side	Yes, No
	Comes with integrated LED light curing unit	Yes, No
	No of programs in integrated LED light curing unit	3, 4, 5, more than 5, NA(Integrated LED light curing unit not provided)
	No of eye protection shield supplied	1, 2, 3, 4
	Suction	Compatible with vacuum capacity of 100mbar or less at 250 litres/min, Has high and low suction with autostart and stop function at both surgeon and assistant side, Has facility to ON/OFF suction through chair based foot switch, Has facility for automatic wash and disinfection system for suction system
	Dentist side control	Overhead design/overhead delivery, Underhanging design
	Dentist side control has 6 way syringe	Yes, No
	Dentist side Control has 3 way syringe	Yes, No

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Dentist side control has Provision for brushless micromotor with speed range upto 40000 rpm	Yes (fibre optic), Yes (non-fibre optic), NA
	Dentist side control has Provision for Fibre optic airotor terminal	1, 2, 3, None
	Dentist side control has provision of Non-Fibre optic airotor terminal	1, 2, 3, None
	Dentist side control has Provision for scalar terminal	Yes (Fibre Optic), Yes (non-fibre optic), NA, Inbuilt
	Controls for airotor and micromotor	On foot rest only, both on control panel and foot rest
	Has provision to automatically emit an airjet when operator stops handpieces so as to clean any residual liquids or solids from the hand piece	Yes, No
	Compressor	Dental Chair shall be supplied with suitable oil free, Medical grade, noise free and with anti-bacterial coating Air Compressor, with capacity to run single dental chair unit, Dental Chair shall be supplied with suitable oil free, Medical grade, noise free and with anti-bacterial coating Air Compressor, with capacity to run up to three dental chair units simultaneously, Dental Chair with out Air Compressor unit
	Tank capacity of compressor	less than 10 litres, 10-20 litres, 20- 30 litres, more than 30 litres
	Integrated back-lit radiographic image viewer	None, For OPG and IOPA, For IOPA only
	Operator stools supplied with chair	1, 2, 3
	No of castors in operator stool	4, 5, 6, more than 6
	Operator stool is with backrest	Yes, No
	Operator stool design	Saddle style stools, basic with adjustable backrest, adjustable stool with armrests and excellent lumbar support, ergonomically contoured seats
	Micromotor (Autoclavable) in built with chair	Yes, No
	Micromotor speed control minimum	30000, 40000, 50000, 55000, 35000
	Ultrasonic Scaler supplied	Yes, No

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Make of Ultrasonic scaler (Autoclavable)	Fibre optic, Non fibre optic, Inbuilt
	Fibre optic contrangle handpiece autoclavable head for micromotor (WITH PUSH BUTTON)	1, 2, 3, NA
	Non Fibre optic contrangle handpiece autoclavable head for micromotor (WITH PUSH BUTTON)	1, 2, 3, NA
	Fibre optic contrangle handpiece autoclavable head for micromotor (WITHOUT PUSH BUTTON)	1, 2, 3, NA
	Non Fibre optic contrangle handpiece autoclavable head for micromotor (WITHOUT PUSH BUTTON)	1, 2, 3, NA
	Autoclavable Fibre optic airotor handpiece (standard head with push button)	1, 2, 3, NA
	Autoclavable Fibre optic airotor handpiece (mini head with push button)	1, 2, 3, NA
	Autoclavable Non Fibre optic airotor handpiece (standard head with push button)	1, 2, 3, NA
	Autoclavable Non Fibre optic airotor handpiece (standard head without push button)	1, 2, 3, NA
	Autoclavable Non Fibre optic airotor handpiece (mini head with push button)	1, 2, 3, NA
	Autoclavable Non Fibre optic airotor handpiece (mini head without push button)	1, 2, 3, NA
	Fibre optic straight handpiece for micromotor (autoclavable)	1, 2, 3, NA
	Non Fibre optic straight handpiece for micromotor (autoclavable)	1, 2, 3, NA
	Periodontal function tips along with scaler	4, 6, 8, More than 8

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Stabilizer	No external stabilizer required (comes with in-built stabilizer), Comes with stand alone stabilizer, No stabilizer provided
	Capacity of stabilizer	2kva, 2.5 kva, 3kva, 4kva, NA
	The main equipment and all accessories/attachments are from reputed OEM's with the product /products displayed in OEM website	Yes, NA
	Power supply	220-230V , 50 Hz single phase
	Warranty in years	3, 4, 5
	CMC offered post warranty in years	3, 4, 5, None
	Whether warranty provided shall be applicable if product used outside India by any buyer	Yes, No
	Control panel	basic control panel with 3 digit display, with LED/LCD display, with Touch panel, Accepts voice commands, without display
	Control panel has Modes available to control airtor and micromotor RPM and torque	Yes, No
	Remote control	Yes (for chair movement, LED light control, spittoon water control, rinsing water control etc), No remote control panel
Material Parameter	Padding	Fixed, Removable
	Padding material	Memory foam, Memory foam with gel filled layer, latex foam, high density PU foam
Additional Parameters	There is possibility to bypass mains water with supply from fresh water bottle	Yes, No
	Comes with integrated medical grade touch screen flat panel mounted on an arm that should rotate and tilt for comfort viewing of doctor and patient	Yes, No
	Size of touch screen flat panel of integrated medical grade screen in inches	18" or less, 18-20", 21-24", More than 24", NA (for not provided)
	OEM representation in Indian market (in Years) hint: Documents should be supplied on buyers demand	1-2 years, 3 years or more, No representation in Indian market

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	The chair might be required to be installed at remote/ field locations in India, OEM will provide prompt Installation, user training, service anywhere in India as and when required	Yes, No
	Product instruction manual to be supplied	Yes, No
	Resolution of screen	1920 x 1080, 1920 x 1200, 1360 x 768
	Comes with in built Intraoral Camera	Yes, No
	Comes with inbuilt RVG	Yes, No

Additional Specification Parameters - Advanced Dental Chair (3 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Specifications attached in ATC are final	Specifications attached in ATC are final only quote as per attached specifications.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Prabhat Sharma	184120,AIIMS VIJAYPUR JAMMU	3	60

Special terms and conditions-Version:3 effective from 24-07-2023 for category Advanced Dental Chair

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license,

- product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
 4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
 5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

 6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

 7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

 8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.

 9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.

 10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.

11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.

14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---