



F-No -AIIMS/JMU/BOO/2024/LP/

Dated 4 June, 2024

Inviting Quotations for Purchase of Item for Paediatrics Department, AIIMS, Vijaypur Jammu

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of items for **Paediatrics Department, AIIMS, Vijaypur, Jammu**. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at **AIIMS, Vijaypur, District Samba, Jammu (184120)** up to **11-06-2024** till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of item are given as under:-

S. No.	Name of Item	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
A.	Self-Inflating bag 200 to 300 ml	1					
B.	Self-Inflating bag 500 ml	1					
C.	Self-inflating bag 750 ml	1					
D.	Self-Inflating bag 1000-1500 ml	1					
E.	Self-Inflating bag 1400-2000 ml	1					

Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. Supply should be made within 15 days from the date of purchase order.
5. Price should be for Destination basis (i.e., concerned department).
6. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
7. **Quotation Name and No. must be mentioned on top of envelope.**
8. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
9. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
10. Sealed quotation should be submitted by speed post/copy courier to Procurement Section, **AIIMS, Vijaypur, District Samba, Jammu (184120)** up to **11-06-2024** till **03:00 pm**.
11. **Validity of quotation should be 90 days from the date of opening.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

12. Sample to be submitted as and when required by the institute.
13. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
14. All column given in quotation should be filled otherwise quotation will not be accepted.
15. Along with quotation please mentioned Email-id and Contact Number.

Takshakundal

Chairperson,
Store Purchase Committee,
AIIMS, Vijaypur, Jammu

ANNEXURE 2

Technical specifications

SPECIFICATIONS		
S.No	Item	Specifications
1.	Self-inflating bag 200 to 300 ml	<p>Self-inflating bag, Silicone made, Provided with closed ended reservoir with two valves. Standard 15/22 mm Swivel connector allows connections to all common masks Endotracheal Tubes both for adults and infants. Patient valves pliable, well-sealed, have minimum dead space and no forward or backward leaks, The bag should have an oxygen inlet which fits into the standard oxygen tubing both from a cylinder and central supply Face masks should be transparent, soft, cushioned, fit the patient outlet easily and have minimum dead space, made of silicon material Face masks: 3 sizes i.e. 00,01,02,: 3 set with each bag. Should have option to control PEEP & PIP Should have pressure release valve at 40 cm H2O. The reservoir should be a PVC bag of 600 ml capacity. Accessories required: Silicon bellow, Non Rebreathing Valve, 2 meter oxygen tube, oxygen reservoir bag (1 extra) Must be supplied in carrying case The resuscitator, its components and all its accessories should be free of natural rubber latex Manufacturer / supplier should have ISO certificate for quality standard. Should be FDA (US) / CE (EU) approved product or BIS certified ; warranty 1 year</p>
2.	Self-inflating bag 500 ml	<p>Self-inflating bag, Silicone made, Provided with closed ended reservoir with two valves. Standard 15/22 mm Swivel connector allows connections to all common masks Endotracheal Tubes both for adults and infants. Patient valves pliable, well-sealed, have minimum dead space and no forward or backward leaks, The bag should have an oxygen inlet which fits into the standard oxygen tubing both from a cylinder and central supply Face masks should be transparent, soft, cushioned, fit the patient outlet easily and have minimum dead space, made of silicon material Face masks: 3 sizes i.e. 00,01,02,: 3 set with each bag. Should have option to control PEEP & PIP Should have pressure release valve at 40cm H2O.</p>

Santhi

Santhi

	<p><i>warranty 1 year</i></p>	<p>The reservoir should be a PVC bag of 600ml capacity. Accessories required: Silicon bellow, Non Rebreathing Valve, 2 meter oxygen tube, oxygen reservoir bag (1 extra) Must be supplied in carrying case The resuscitator, its components and all its accessories should be free of natural rubber latex Manufacturer / supplier should have ISO certificate for quality standard. Should be FDA (US) / CE (EU) approved product or BIS certified</p>
<p>3.</p>	<p>Self-inflating bag 750 ml</p>	<p>Self-inflating bag, Silicone made, Provided with closed ended reservoir with two valves. Standard 15/22 mm Swivel connector allows connections to all common masks Endotracheal Tubes both for adults and infants. Patient valves pliable, well-sealed, have minimum dead space and no forward or backward leaks, The bag should have an oxygen inlet which fits into the standard oxygen tubing both from a cylinder and central supply Face masks should be transparent, soft, cushioned, fit the patient outlet easily and have minimum dead space, made of silicon material Face masks: 3 sizes i.e. 00,01,02,: 3 set with each bag. Should have option to control PEEP & PIP Should have pressure release valve at 40cm H2O. The reservoir should be a PVC bag of 1000 ml capacity. Accessories required: Silicon bellow, Non Rebreathing Valve, 2 meter oxygen tube, oxygen reservoir bag (1 extra) Must be supplied in carrying case The resuscitator, its components and all its accessories should be free of natural rubber latex Manufacturer / supplier should have ISO certificate for quality standard. Should be FDA (US) / CE (EU) approved product or BIS certified</p>
<p>4.</p>	<p>Self inflating bag 1000-1500 ml</p>	<p>Self-inflating bag, Silicone made, Provided with closed ended reservoir with two valves. Standard 15/22 mm Swivel connector allows connections to all common masks Endotracheal Tubes both for adults and infants. Patient valves pliable, well-sealed, have minimum dead space and no forward or backward leaks, The bag should have an oxygen inlet which fits into the standard oxygen tubing both from a cylinder and central supply Face masks should be transparent, soft, cushioned, fit the patient outlet easily and have minimum dead space, made of silicon material Face masks: 2 sizes i.e. 03, 04: 3 set with each bag.</p>

warranty 1 year

GA
Amish

		<p>Should have option to control PEEP & PIP</p> <p>Should have pressure release valve at 40cm H2O.</p> <p>The reservoir should be a PVC bag of 1500-2600 ml capacity.</p> <p>Accessories required: Silicon bellow, Non Rebreathing Valve, 2 meter oxygen tube, oxygen reservoir bag (1 extra)</p> <p>Must be supplied in carrying case</p> <p>The resuscitator, its components and all its accessories should be free of natural rubber latex</p> <p>Manufacturer / supplier should have ISO certificate for quality standard. Should be FDA (US) / CE (EU) approved product or BIS certified <i>warranty 1 year</i></p>
5.	Self inflating bag 1400-2000 ml	<p>Self-inflating bag, Silicone made,</p> <p>Provided with closed ended reservoir with two valves.</p> <p>Standard 15/22 mm Swivel connector allows connections to all common masks Endotracheal Tubes both for adults and infants.</p> <p>Patient valves pliable, well-sealed, have minimum dead space and no forward or backward leaks,</p> <p>The bag should have an oxygen inlet which fits into the standard oxygen tubing both from a cylinder and central supply</p> <p>Face masks should be transparent, soft, cushioned, fit the patient outlet easily and have minimum dead space, made of silicon material</p> <p>Face masks: 3 sizes i.e. 03, 04, 05: 3 set with each bag.</p> <p>Should have option to control PEEP & PIP</p> <p>Should have pressure release valve at 40cm H2O.</p> <p>The reservoir should be a PVC bag of 1500-2600 ml capacity.</p> <p>Accessories required: Silicon bellow, Non Rebreathing Valve, 2 meter oxygen tube, oxygen reservoir bag (1 extra)</p> <p>Must be supplied in carrying case</p> <p>The resuscitator, its components and all its accessories should be free of natural rubber latex</p> <p>Manufacturer / supplier should have ISO certificate for quality standard. Should be FDA (US) / CE (EU) approved product or BIS certified <i>warranty 1 year</i></p>

- warranty - 1 year for all systems
Shreey
07/05/2024

Sankh

Q