



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

F-No.-AIIMS/JMU/BOO/2024/LP/

Dated: 13 June, 2024

Inviting Quotations for Purchase of Instrument for **Otorhinolaryngology & Head Neck Surgery (ENT OPD)** Department, AIIMS, Vijaypur Jammu.

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of Instrument for **Otorhinolaryngology & Head Neck Surgery (ENT OPD) Department**, AIIMS, Vijaypur, Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, **Vijaypur, District Samba, Jammu (184120)** up to **2006-2024** till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of item are given as under:-

S. No.	Name of Instrument	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1.	Otoscope (LED source with USB – port charger-rechargeable)-Heine/welchallyn	01					
2.	Tongue depressor (set of 3)	01 set					
3.	Nasal speculum-Thudicum (set of 3)	01 set					
4.	Aural speculum (set of 4)	01 set					
5.	Tilleys nasal forceps	01					
6.	Aural forceps(micro cup forceps) 14 cm	01					
7.	Aural forceps(micro crocodile forceps) 14 cm	01					
8.	Jobsons wax probe	01					
9.	Iris scissors(Small)	01					
10.	Adsons toothed forceps, 12.1 cm	01					
11.	Adsons nontooth forceps, 12.1 cm	01					
12.	Suction tips zollner's(Set of 4)	01 set					
13.	Ear suction tips with adaptor (set of 5)	01 set					



\*\*\*\*\*

14.	I/L mirror with handle (set of 4) (autoclavable)	01 set					
15.	P/R mirror with handle (set of 4) (autoclavable)-	01 set					
16.	Spirit lamp with spirit bottle	01					
17.	Eustachian Tube catheter	01					
18.	Punch Biopsy forceps (straight, 45-degree, right angle- set of one each), 17.5 cm	01					
19.	Kidney tray, 12 inch	01					
20.	Instrument tray, 12x8 Inch	01					
21.	Instrument trolley, 20x30 Inch	01					
22.	Killian's speculum	01					
23.	B.P. handle	01					
24.	Bite block	01					
25.	Cheatle's Forceps	01					
26.	Straight artery forceps (6 inches and 4 inches)	01 each					
27.	Curved artery forceps (6 inches and 4 inches)	01 each					
28.	Toothed forceps 6"	01					
29.	Nontoothed forceps 6"	01					
30.	Tuning Fork (256,512,1024 Hz- set)	01 set					
31.	Suction Machine (Double Jar)	01					
32.	Tracheostomy Set						
a)	BP Handle No. 3	01					
b)	Fine mosquito artery forceps curved (4 inches)	01					
c)	Fine mosquito artery forceps curved (6 inches)	01					
d)	Langen beck Retractor (small)	01					
e)	Langen beck Retractor (medium)	01					
f)	Tracheal dilator forceps (4.75 inches)	01					



\*\*\*\*\*

g)	Cricoid Hook (single prong, blunt) (6.25 inches)	01					
h)	Tracheal Hook (single prong, sharp) (6.25 inches)	01					
i)	Tracheal Hook (double prong, blunt) (6.25 inches)	01					
j)	Allis Tissue Forceps small (5 Inches)	01					
k)	Needle holder small (6 inches)	01					
l)	Sponge holding Forceps small (8 inches)	01					
m)	Kidney tray small (10 inches)	01					
n)	Bowel (medium)( 13x9x2.5 inches)	01					
o)	Straight artery forceps (4 inches)	01					
p)	Straight artery forceps (6 inches)	01					
q)	Adson Tissue Holding forceps (toothed) (5 inches)	01					
r)	Adson Tissue Holding forceps (non- toothed) (5 inches)	01					
33.	Metallic drum(12X15 inch) or larger	01					
34.	Rechargeable Head Lights	01					
35.	Yankauer suction set	01					

### Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. Supply should be made within 15 days from the date of purchase order.
5. Price should be for Destination basis (i.e., concerned department).
6. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
7. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

8. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
9. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
10. Sealed quotation should be submitted by speed post/copy courier to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to 20-06-2024 till 03:00 pm.
11. Validity of quotation should be 90 days from the date of opening.
12. Sample to be submitted as and when required by the institute.
13. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
14. All column given in quotation should be filled otherwise quotation will not be accepted.
15. Along with quotation please mentioned Email-id and Contact Number.

*Rakshakundal*  
13/June 2024  
Dr Raksha Kundal  
Chairperson,  
Store Purchase Committee,  
AIIMS, Vijaypur, Jammu

**Department of Otorhinolaryngology & Head Neck Surgery**

**List of instruments for ENTOPD**

<b>S.No.</b>	<b>Name Of the Instrument</b>
1.	Otoscope (LED source with USB port charger-rechargeable)-Heine/welchallyn
2.	Tongue depressor
3.	Nasal speculum- Thudicum
4.	Aural speculum
5.	Tilleys nasal forceps
6.	Aural forceps(micro cup forceps) 14 cm
7.	Aural forceps(micro crocodile forceps) 14 cm
8.	Jobsons wax probe
9.	Iris scissors(Small)
10.	Adsons toothed forceps, 12.1 cm
11.	Adsons nontooth forceps, 12.1 cm
12.	Suction tips zollner's
13.	Ear suction tips with adaptor
14.	I/L mirror with handle (autoclavable)
15.	P/R mirror with handle (autoclavable)
16.	Spirit lamp with spirit bottle
17.	Eustachian Tube catheter
18.	Punch Biopsy forceps (straight, 45-degree, right angle- set of one each), 17.5 cm
19.	Kidney tray, 12 inch
20.	Instrument tray, 12x8 Inch
21.	Instrument trolley, 20x30 Inch
22.	Killian's speculum
23.	B.P. handle
24.	Bite block
25.	Cheate's Forceps
26.	Straight artery forceps (6 inches and 4 inches)
27.	Curved artery forceps (6 inches and 4 inches)
28.	Toothed forceps 6"
29.	Nontoothed forceps 6"
30.	Tuning Fork (256,512,1024 Hz- set)
31.	Suction Machine (Double Jar)
32.	Tracheostomy Set
33.	Metallie drum(12X15 inch) or larger
34.	Rechargeable Head Lights
35.	Yankauer suction set

### Otoscope (1 in Number)

- Light Source: LED/3.5V Halogen. Otoscope Set with rechargeable handle/ Battery
- Magnification: 4.2x or 3x
- With an all metal housing and a scratch resistant glass window
- Insufflation port for pneumatic mobility testing (pneumatic otoscopy)
- Otoscope LED, with 1 set (4 pcs.) of reusable tips (B-000.11.111), 10 All Spec disposable tips 4 mm , hard case, rechargeable handle with table charger
- LED in Quality – or LED HQ (Fibre Optic Illumination. Ensures homogeneous, very bright illumination and an unobstructed view of the ear canal and tympanum). Li-ion technology, rapid charging and an additional charge status indicator
- Exclusive continuous brightness control between 100% and 3%
- Optimal light intensity can be individually adjusted
- Hard case
- 5 Year Guarantee

### Suction Machine

- High Vacuum suction machine -Medical (Aspiration unit)
- Material of cabinet (Base, top and panel) - Epoxy powder coated
- Mild Steel (MS) with SS Top
- Should have Rust proof housing cabinet and castor wheels
- The unit should be mounted in such a way as to produce minimum vibrations
- Material of twin jars/bottles - unbreakable Polycarbonate (PC)
- Suction bottles/jars with Synthetic rubber lid and fitted with arrangement to prevent overflow of fluid to outside
- Capacity of each suction bottle/jar graduated – 2000ml
- Interconnected jar/bottle with overflow protection device
- Double overflow protection with safety device provided
- Non-collapsible suction tubing ID -10 mm
- Material of suction tube- Silicon
- Minimum Length of suction tube – 2.4 meter
- Should have Flutter free vacuum control knob
- Bacterial suction filter/air filters and water traps between the suction bottle and suction pump provided
- Material of tray- Stainless Steel (SS)

- Vacuum Gauge (Bourdon tube type) with concentric scale marking in black on white background, mounted in panel, easily visible
- During use of machine &quot;7.5cm dia, 0 to 760mmHg&quot;
- Entire unit mounted on four castor wheels with minimum two break
- Material of castor wheels should be Thermoplastic Rubber (TPR)
- Facility for easy maintenance - Back open
- Size (HxWxL) (maximum) - 750X300X330
- Weight (Kg) - 7 to 12
- OPERATIONAL & TECHNICAL REQUIREMENTS
- input power supply 220+/-10% V, 50Hz, Single phase AC and also
- 12 V, DC battery operated.
- Power consumption (Watt)- 380
- Power indicator should be present.
- Integrated on/off switch on the trolley should be present.
- Motor capacity - 0.5 (HP)
- Type of suction pump - Oil-free
- Heavy duty motorized pump for fast vacuum built-up Yes
- Minimum air flow displacement of suction pump (Ltrs/Minute)
- to 50
- Built-up Vacuum (Maximum) &quot;-713 +/-10mmHg (-95kPa)&quot;
- Sound level (Maximum) (dB)
- STANDARD ACCESSORIES
  - Compatible 3 core power cord with Indian plug- 5-meter length
  - Jars/ Bottles
  - Suction tubing
- USFDA/European CE certified/ISI marked
- Onsite installation /services to be provided by the seller
- 2-year warranty

#### Tracheostomy set instruments:

S. No.	Instrument name
1	BP Handle
2	Fine mosquito artery forceps curved (4 inches)
3	Fine mosquito artery forceps curved (6 inches)
4	Langen beck Retractor (small)

5.	Langen beck Retractor (medium)
6.	Tracheal dilator forceps (4.75 inches)
7.	Cricoid Hook (single prong, blunt) (6.25 inches)
8.	Tracheal Hook (single prong, sharp) (6.25 inches)
9.	Tracheal Hook (double prong, blunt) (6.25 inches)
10.	Allis Tissue Forceps small (5 inches)
11.	Needle holder small (6 inches)
12.	Sponge holding Forceps small (8 inches)
13.	Kidney tray small (10 inches)
14.	Bowel (medium)( 13x9x2.5 inches)
15.	Straight artery forceps (4 inches)
16.	Straight artery forceps (6 inches)
17.	Wilson Tissue Holding forceps (toothed) (5 inches)
18.	Wilson Tissue Holding forceps (non-toothed) (5 inches)
	<b>Total</b>

- Sterilizable by Autoclaving and Chemical sterilization.
- The OPD metallic instruments must be made up of high quality medical grade hardened stainless steel with defined specifications like AISI-410, AISI-420, AISI-304, AISI-303, AISI-440 etc. using guidelines of ASTM Standard F899-94 and ISO7153 and with a dull finish.
- The instrument should be non-reflective.
- The instrument should not be a must (Video physical)
- The instrument should have locking catches.
- The instrument should have a ratchet mechanism.
- The instrument should have a straight is already mentioned in list of articles should be included.



## Rechargeable Head Light

- High Power LED
- 10 W high power LED emit up to 1,10,000 Lux illuminance
- Ergonomic Design
- Detachable forehead cushion and multiple adjustment headband
- Integration Design
- 100% freedom of movement with cable free intergrated battery
- High CRI Value of LED
- Superb Color Quality with High Contrast
- Adjustable Light Spot Size
- Sharp Focus of Light Spot.
- Zoom Knob to adjust the size of the light spot.
- SPECIFICATIONS
  - High Power LED :- 10 Watt
  - Mini portable wireless design
  - High Illumination :- 1,10,000 lux
  - Battery :- Rechargeable Li-Ion
  - Bulb Life :- 50000 Hours
  - CRI  $\leq$  90
  - Color Temperature :- 5700k
  - Working Hour per charge:- at least 4 Hours
  - Charging time :- 1.5 Hour
  - Light Intensity :- Adjustable
  - Light Spot Size :- Adjustable
  - Voltage :- DC3.7 Volt
  - Extra Battery Optional

## Instruments

- Sterilizable by Autoclaving and Chemical sterilization.
- The OPD metallic instruments must be made up of high quality medical grade hardened stainless steel with defined specifications like AISI-410, AISI-420, AISI-304, AISI-303, AISI-440 etc. using guidelines of ASTM Standard F899-94 and ISO7153 and with a dull finish.
- The instruments should be non-reflective.
- The instrument demonstration is a must (Video/physical).