

ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIIJAYPUR, JAMMU-184120 (A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

F-No.-AIIMS/JMU/BOO/2024/LP/

Dated: 25 June, 2024

Inviting Quotations for Purchase of Instrument for Pain Clinic and Surgical Gastroenterology Department, AIIMS, Vijaypur Jammu.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of Instrument for Pain Clinic and Surgical Gastroenterology Department, AllMS, Vijaypur, Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AllMS, Vijaypur, District Samba, Jammu (184120) up to \$\mathcal{g}\$-07-2024 till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of item are given as under:-

S. No.	Name of Instrument	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
Pain	Clinic					1	
1.	Surgical Steel Bowl (80 x 40 mm)	01					,
2.	Surgical Steel Bowl (116 x 50 mm)	01					
3.	Surgical Steel Bowl (160 x 65 mm)	01					
4.	Surgical Steel Bowl (147 x 65 mm)	01					
Surgi	cal Gastroenterology			•			
1.	Hemorhoidal banding set	01				,	
2.	Rigid Sigmoidoscope	01					
3.	Minor OT Set						
a)	Surgical Steel Bowl (80 x 40 mm)	01					
b)	Surgical Steel Bowl (116 x 50 mm)	01					
c)	Lister Bandage cutting Scissor 8 inches	01					
d)	Bozemann/Wertheim straight needle holder (Tungsten Carbide tip) 22-24 cms	01					
e)	Crile wood straight needle holder 16-18 cm	01					
f)	Forster Sponge Holder, Straight Smooth jaws 18 cm	01					



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIIJAYPUR, JAMMU-184120 (A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

Dissecting forceps toothed 14.5 cm	01	
Dissecting forceps non-toothed 14.5 cm	01	
Clip Remover	01	
	Dissecting forceps non-toothed 14.5 cm	Dissecting forceps non-toothed 14.5 cm

Terms & Condition

- 1. Firm to mention Make/Brand name in their quotation.
- 2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
- 3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
- 4. GST/other taxes please may be included in your quotation with the price in separate column.
- 5. Supply should be made within 15 days from the date of purchase order.
- 6. Price should be for Destination basis (i.e., concerned department).
- 7. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
- 8. Quotation Name, Name of Department and No. must be mentioned on top of envelope.
- 9. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
- 10. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
- 11. Sealed quotation should be submitted by speed post/copy courier to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to §-07-2024 till 03:00 pm.
- 12. Validity of quotation should be 90 days from the date of opening.
- 13. Sample to be submitted as and when required by the institute.
- 14. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
- 15. All column given in quotation should be filled otherwise quotation will not be accepted.
- 16. Along with quotation please mentioned Email-id and Contact Number.

Dr Raksha Kundal Chairperson,

Store Purchase Committee, AIIMS, Vijaypur, Jammu

INSTRUMENTS FOR PAIN CLINIC

· · ·	
-	
u ,	
	-

GENERAL SPECIFICATIONS OF THE ABOVE-MENTIONED EQUIPMENT

The Instruments quoted should be of high quality and standard which should meet the international standards.

The Instruments should have European CE or USFDA certification.

Copy of the European CE certificate or USFDA certificate should be enclosed.

The Instruments must be ISO certified ISO 7153 and copy should be enclosed.

The company should provide material certificates.

Instruments should carry Laser markings indicating the Manufacturer's name and Batch No. or Lot No.

Sterilization container should be quoted along with instrument set.

The sterilization containers should meet International standards and approved for steam sterilization procedures to EN 285: 2008 and validated according to ISO 17665 Part: 2006.

Sterilization container and instruments should be of the same parent company

Physical demonstration should be arranged at the time of requirement.

Each instrument should have the following information printed on it: Article No., Manufacturer's Name, Country of Origin, Batch/Lot No.

A bar code / QR code that should give the following information for easy stock management:

a). Part No. and Instrument name b). Date of Production / Lot No.

Warranty for all items: 3 years

SPECIFICATIONS FOR HEMORRHOIDAL BANDING SET

It should be made up of high quality surgical grade stainless steel.

It should be supplied with a stepped cylindrical cone (Ø 10mm) for easy band loading, or with a smooth cone and a ring-loader.

It should be provided with a grasping forceps and an anoscope.

A set of 100 rings to be provided along

It should have ergonomic design for optimal visual clarity.

The Instruments must be ISO certified ISO 7153 and copy should be enclosed.

Bidder must enclose original literatures & technical data sheet in the support of the technical bid.

Company should provide material certificates.

Instruments should carry Ledger marking indicating the Manufacturer's name and Batch No. or Lot No.

Instrument should be autoclavable.

Each instrument should have the following information printed on it: Article No., Manufacturer's Name, Country of Origin, Batch/Lot No.

A bar code / QR code that should give the following information for easy stock management: a). Part No. and Instrument name b). Date of Production / Lot No.

Minimum warranty period of 3 years

Reusable Rigid Sigmoidoscopes/Anoscopes:

1. Standard Sigmoidoscopes

19 mm x 15 cm Standard Sigmoidoscope, Speculum only with Obturator 19 mm x 25 cm Standard Sigmoidoscope, Speculum only with Obturator 19 mm x 30 cm Standard Sigmoidoscope, Speculum only with Obturator 19 mm x 35 cm Standard Sigmoidoscope, Speculum only with Obturator

23 mm x 25 cm Standard Sigmoidoscope, Speculum only with Obturator

b. Complete Illumination System

Light Handle with 0.9 m Cord

Light Handle plus Transformer

Fiber Optic Light Head

Transformer with 1.5 m Cord

3.5 V Halogen Handle Adapter

A 3.5 V Direct Plug-in Rechargeable Handle.

Insufflation Bulb, complete

Hydrophobic Filters (pack of 10) with Luer Loc

- 2. Pediatric/Stricture Sigmoidoscopes 32010
 - 11 mm x 10 cm Fiber Optic/Stricture Speculum only with Obturator
 - 11 mm x 25 cm Fiber Optic/Stricture Speculum only with Obturator
 - 15 mm x 15 cm Fiber Optic/Pediatric Speculum only with Obturator
 - 15 mm x 25 cm Pediatric/Stricture Speculum only with Obturator
- -3. Rotating Anoscope

Anoscope Only 68 mm Speculum Length
Rotating Speculum Anoscope with Obturator, 20 mm inside diameter, 22 mm outside diameter

4. Fiber Optic Anoscope Speculum Only with Obturator

19 mm x 7 cm Fiber Optic Anoscope, Speculum Only with Obturator

23 mm x 7 cm Fiber Optic Anoscope, Speculum Only with Obturator

27 mm x 7 cm Fiber Optic Anoscope, Speculum Only with Obturator

Anoscope Assesories:

Light Handle with 0.9 m Cord Light Handle plus Transformer Fiber Optic Light Head Transformer with 1.5 m Cord :Scopes

- •All SurgicalInstruments are European CE approved and US-FDA registered.
- •Physical demonstration is mandatory be provided.
- •Material used in surgical instruments should be German Steel conforming to DIN 58298:2010-09 " Medical instruments-Materials, finish and testing"
- •All steel instruments should be guaranteed against metal fatigue and rust.
- •All surgical steel instruments should be made of high grade surgical stainless steel 420 grade, autoclavable.
- •Brand name, catalogue number and Data Matrix UID should be mentioned on each instrument.
- •All our instruments should be easy to use, extremely user-friendly and ergonomically designed
- •All Instruments must be from Single manufacturer.

· Should have Sy warranty + Sy CAMC

Department of Surgical Gastroenterology

- 1. Surgical steel bowl 80*40 mm -
- 2. Surgical steel bowl 116*50 mm/
- 3. Lister bandage cutting scissor 8-inch
- 4. Wertheimstraight needle holder Vungilen -carlide top 2272464
- 5. Crile wood straight needle holder 16 -118 cm
- 6. Forster sponge holder / 18 e m
- 7. Dissecting forceps toothed 14.5 cm ~
- 8. Dissecting forceps non-toothed 14.5 cm
- 9. Clip remover

All Surgical Instruments should be European CE approved and US-FDA registered.

- The material testing certificate should be enclosed in the technical bid.
- Material used in surgical instruments should be German Steel conforming to DIN 58298:2010-09, Medical materials, finish and testing.
- All instruments should be guaranteed against metal fatigue and rust.
- All surgical instruments should be made of high-grade surgical stainless steel 420 grade, autoclavable.
- Brand name, catalogue number and Data Matrix UID should be mentioned on each instrument.
- All Surgical Instruments should be capable of being used on the most fragile tissues and finest sutures.
- All our instruments should be easy to use, extremely user-friendly and ergonomically designed.
- All Instruments must be from a single manufacturer.
- Should 3-year warranty