



\*\*\*\*\*

F-No. -AIIMS/JMU/BOO/2024/LP/01

Dated: 19 July, 2024

Inviting Quotations for Purchase of items required for Anaesthesiology Department, AIIMS, Vijaypur Jammu.

### QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of items for **Anaesthesiology** Department, AIIMS, Vijaypur, Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, Vijaypur, District Samba, Jammu (184120) up to **26-07-2024** till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of item are given as under:-

| S. No. | Name of Instrument                             | Qty. | Packs Size/Unit | HSN Code | Basic Price | GST % | Total Cost Inclusive of GST |
|--------|--|------|-----------------|----------|-------------|-------|-----------------------------|
| 1      | Monitor with protective case                   | 01   |                 |          |             |       |                             |
| 2      | Video-Flexible Bronchoscope Size: 1.0-1.2 mmID | 01   |                 |          |             |       |                             |
| 3      | Video-Flexible Bronchoscope Size: 3.6-3.8 mmID | 01   |                 |          |             |       |                             |

### **Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. Supply should be made within 15 days from the date of purchase order.
6. Price should be for Destination basis (i.e., concerned department).
7. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
8. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
9. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
10. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
11. Sealed quotation should be submitted by speed post/copy courier / by hand to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to **26-07-2024** till 03:00 pm.



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

12. Validity of quotation should be 90 days from the date of opening.
13. Sample to be submitted as and when required by the institute.
14. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
15. All column given in quotation should be filled otherwise quotation will not be accepted.
16. Along with quotation please mentioned Email-id and Contact Number.

*Rakshakundal*  
*19/7/24*

**Dr Raksha Kundal**  
**Chairperson,**  
**Store Purchase Committee,**  
**AIIMS, Vijaypur, Jammu**

## Video Bronchoscope Pre-Sterile Disposable Flexible Bronchoscope System

### Specifications of Monitor = Protective Case

- It should be Lightweight and easy to set up on patient table / IV Pole
- It should have minimum internal storage capacity of 32 GB
- It should have inbuilt rechargeable Lithium-ion Battery with 3 Hours back-up
- It should have 2 USB Interface ports to transfer recorded High-resolution video and still images
- Monitor should have HDMI cable for connecting to a big screen
- The screen size should be minimum 12.5" (Diagonal)
- The Screen should have Image Extend feature for enhanced viewing of Live images
- It should display full frame HD resolution view and no circular view
- High-resolution video and still images can be recorded with touch on the screen
- It should have wi-fi capability to transfer data viz PACS and DICOM
- Capability to attach two scopes at a time with split screen facility.
- Flexible scopes should be connected directly with monitor without any special accessory
- Scopes & monitor should be from same manufacturer
- It should be capable of easy navigation and fast identification of anatomical landmarks
- The scopes should be sterile packed & minimum length of the scope should be 600mm & 300mm.
- It should have ET Tube parking slot & should have Intermittent Suction Controller button to control suction on pressing
- The working channel should allow fluid instillation and oxygen flushing
- An Introducer should be provided to facilitate introduction of endoscopic accessories with each scope.
- Should have Control Lever on handle for the movement of distal tip up & Down
- **Should be supplied with single use scopes for MDR, Immune Compromise, Positive, Covid-19 and other fungal infection.**
- **Should supplied with one ETO Pre Sterile closed-loop system for BAL sampling with scopes.**
- Flexible scopes should be suitable for the following procedures:
  - ❖ Intubation (oral & nasal)
  - ❖ Bed side Bronchoscopy
  - ❖ Airway Inspection
  - ❖ Management of retained secretion
  - ❖ BAL & BW procedures
  - ❖ DLT placement and checking
  - ❖ PDT procedures

### Technical Features of Scope

| Features       | Slim Scope                         | Adult Scope                        |
|----------------|------------------------------------|------------------------------------|
| Inner Diameter | 1.0 - 1.2 mm                       | 2.0 - 2.2 mm                       |
| Outer Diameter | 3.6 mm- 3.8 mm                     | 4.8- 5.0 mm                        |
| Working Length | Minimum 600 mm                     | Minimum 600 mm                     |
| Bending Range  | Minimum 170° up, minimum 170° down | Minimum 170° up, minimum 170° down |
| Field of view  | 80-85°                             | 80-85°                             |
| Depth of Field | 6 - 50 mm                          | 6 - 50 mm                          |

The price of the consumables has to be fixed for next 05 years from the date of installation of main equipment and to be quoted separately. Should be supplied with ETO Pre Sterile closed-loop system For BAL sampling.

- Should be Single operator sampling solution. Should directly attach to scope with attachment lock
- Should have flow Switch facility to easy switch from suctioning to sampling
- Should have closed lid to provide transportation safe sealing of sample container
- Should have Suction adapter for connection of male style suction tube
- Sample Container should be transparent for easy monitoring of sample volume and texture
- The sample container volume should be 30 ml
- The system should be US-FDA and EU CE approved / ISO / COSCO / BIS
- The whole system supplied as 1 monitor, 1 Broncho sampler and 5 nos of sterile scopes (Scopes sizes as per user choice)
- The rate of consumables should be quoted separately for future purchases.
- It should have ~~one~~ year warranty

(05) Five