F-No.-AIIMS/JMU/BOO/2024/LP/O/

Dated: 24July, 2024

Inviting Quotations for Purchase of OPD Equipment required for Pathology Departments, AIIMS, Vijaypur Jammu.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of OPD Equipment for Pathology Department, AIIMS, Vijaypur, Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, Vijaypur, District Samba, Jammu (184120) up to \$\frac{1}{2}\$-07-2024 till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of item are given as under:-

S. No.	Name of Equipment	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	Semi-automated Coagulometer	01					

Terms & Condition

- 1. Firm to mention Make/Brand name in their quotation.
- 2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
- 3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
- 4. GST/other taxes please may be included in your quotation with the price in separate column.
- 5. Supply should be made within 15 days from the date of purchase order.
- 6. Price should be for Destination basis (i.e., concerned department).
- 7. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
- 8. Quotation Name, Name of Department and No. must be mentioned on top of envelope.
- 9. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
- 10. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
- 11. Sealed quotation should be submitted by speed post/copy courier / by hand to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to 31-07-2024 till 03:00 pm.
- 12. Vendors would get their sample check by user department through procurement section before the last date of quotation opening.
- 13. Validity of quotation should be 90 days from the date of opening.
- 14. Sample to be submitted as and when required by the institute.
- 15. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.





ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIIJAYPUR, JAMMU-184120 (A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

16. All column given in quotation should be filled otherwise quotation will not be accepted.

17. Along with quotation please mentioned Email-id and Contact Number.

Dr Raksha Kundal

Chairperson,

Store Purchase Committee,

1. SEMI-AUTOMATED COAGULATION ANALYSER

- Complete semi-automated coagulometer system with printer is required.
- 2. At least 16 incubation positions for samples
- 3. At least 2 measurement channels.
- 4. At least 2-4 positions for reagents.
- At least 4 independents built in timers for incubation.
- Validated for clotting based coagulation measurements in 3.2% trisodium citrate plasma
- Should have automatic pipetting electronically connected or manual start up
- Should have Backlight LCD display, at least 4 lines of 40 characters with built in printer.
 Should give results in Seconds and in various calculated units (%, INR, Ratio, Gm/L, mg/dL).
- Should have Incubation and measurement wells at 37°C +/- 0.5°C
- Should have Test menu: PT, PTT, TT, FIB (Clauss or TT derived), Factor II, V, VII, VIII, IX, X, XI, XII, Fletcher, VT (Venom time), APCR, AT-III (clot), Protein C (clot), Protein S (clot), Heparin, STAT (PT/PTT)
- 11. The instrument should have ability to save factor level calibration curves and give direct results for coagulation factor levels e.g. Factor VIII etc.
- 12. One complete set of accessories should be provided which should include:
 - Measurement Cuvettes with corresponding magnetic bead/stirrer:1000 Pieces
 - Compatible calibrated micropipette: 1 Piece (25/50/100/200µl)
 - Dispenser for magnetic bead/stirrer: 1 Piece
 - Printer: 1 Piece (not needed additional if in built printer already present).
 - Thermal Paper: 10 rolls
- 13. Data transfer Cable, 1 Pc.
- 14. The unit shall be capable of being stored continuously in ambient temperature of 0 -50deg C and relative humidity of 15-90%
- 15. The unit shall be capable of operating continuously in ambient temperature of 10 -40deg C and relative humidity of 15-90%
- 16. Power input should be 220-240VAC, 50Hz fitted with Indian plug
- Suitable UPS with maintenance free batteries for minimum one-hour back-up should be supplied with the system.
- 18 Comprehensive warranty for 2 years and 5 years AMC after warranty
- 19. Manufacturer should have ISO certification for quality standards.
- 20. Should be FDA or CE or ISI approved product
- 21. Certificate of calibration, inspection from factory and IQ-OQ-PQ must be provided.
- 22. Support Tool kit should be provided for calibration and routine maintenance support as per manufacturer documentation in service / technical manual.
- 23. Log book with instruction for daily, weekly, monthly and quarterly maintenance checklist.