



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

F-No.-AIIMS/JMU/BOO/2024/LP/

Dated: 11 July, 2024

Inviting Quotations for Purchase of Surgical Instruments required in the Operation Theatre for **Trauma and Emergency** Department, AIIMS, Vijaypur Jammu.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of Surgical Instruments for **Trauma and Emergency** Department, AIIMS, Vijaypur, Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, **Vijaypur, District Samba, Jammu (184120)** up to 17-07-2024 till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of item are given as under:-

S. No.	Name of Instrument	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	TC METZENBAUM SCISSORS DEL CVD 180MM	01					
2	TC METZENBAUM SCISSORS DEL CVD 200MM	01					
3	Nerve hook 7"	01					
4	Ob wegeseer soft tissue retractor large	01					
5	Ob wegeseer soft tissue retractor medium	01					
6	Ob wegeseer soft tissue retractor small	01					
7	obwegeseer awl for zygomatic arch	01					
8	Obwegeseer chin retractor	01					
9	Obwegeseer Mandible channel retractor	01					
10	Obwegeseer Mandible rim retractor	01					
11	Obwegeseer Nasal spine retractor	01					
12	Obwegeseer ramus retractor	01					
13	Olivecrona spatula of all sizes	01					
14	Pen suction cannulas all sizes	01					
15	Periosteal elevator big size	01					
16	Periosteal elevator small size	01					
17	Plaster cutting scissor	01					



18	Pool suction cannula	01					
19	Proctoscope speculum large	01					
20	Proctoscope speculum medium	01					
21	Proctoscope speculum small	01					
22	Ribbon retractor	01					
23	Ribshear giertz	01					
24	Richardson retractor	01					
25	Robert's artery forceps	01					
26	Roux Retractor set of 3	01					
27	Ruller stainless steal	01					
28	Ruskin Bone Ronguer	01					
29	Russian tissue forceps 8"	01					
30	Schlesinger Ronger serated	01					
31	self-retainer 6 inch	01					
32	self-retainer hinged 5 inches	01					
33	Self-retaining retractor Balfour child	01					
34	Self-retaining retractor Balfour Adult	01					
35	self-retainer 8 inch	01					
36	semb raspatory	01					
37	Senn Miller retractor D/E SH 16 CM Cats paw	01					
38	Sims speculum set of 3	01					
39	Single hook retractor (tracheal)	01					
40	Skin hooks blunt tip	01					
41	Skin hooks double-sharp	01					
42	Skin hooks sharp tip	01					
43	SL Retractor with double hinge	01					
44	SL RING HDL retractor with double hinge	01					
45	Spatula Malleable 12 x 200mm	01					
46	Sponge holding forceps 10"	01					
47	Sponge-holding forceps 8"	01					
48	Sponge holding forceps 9"	01					
49	SS bowel big & small	01					
50	SS tray with lid	01					
51	Humby Skin Grafting Knife	01					



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

52	DENTRO Bailey Rib Approximator	01					
----	--------------------------------	----	--	--	--	--	--

Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. Supply should be made within 15 days from the date of purchase order.
6. Price should be for Destination basis (i.e., concerned department).
7. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
8. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
9. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
10. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
11. Sealed quotation should be submitted by speed post/copy courier to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to **17-07-2024 till 03:00 pm.** *OR BY HAND*
12. **Validity of quotation should be 90 days from the date of opening.**
13. **Sample to be submitted as and when required by the institute.**
14. **Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.**
15. **All column given in quotation should be filled otherwise quotation will not be accepted.**
16. **Along with quotation please mentioned Email-id and Contact Number.**

Taxshakshy
11/July/2024
Dr Raksha Kundal
Chairperson,
Store Purchase Committee,
AIIMS, Vijaypur, Jammu

Specifications

1. All instruments should be of international quality & and made from surgical-grade stainless steel with carbon bonding on them.
2. All instruments should preferably have Tungsten Carbide Inserts / Carbon Inserts in the Jaws.
3. The Instruments must be vacuum hardened in order to receive the necessary hardness, tenacity, and corrosion resistance. Supporting test reports/certificates must be furnished.
4. The Instruments should have a certificate of Primary passivation process.
5. Every instrument has to be delivered with artificial passivation in the form of a chromium oxide layer to support the natural mutation of this layer from the beginning to make every instrument more corrosion resistant. This passivation process must be executed according to ASTM A967
6. The instrument steel should comply with EN ISO 7153 standard for stainless steel used for surgical instruments and ISO 13402 standards: Surgical hand instruments- determination of resistance against autoclaving, corrosion, and thermal exposure. The necessary certificate should be furnished.
7. The hinges should be rust-proof.
8. The instruments should be guaranteed against metal fatigue and rust.
9. The instrument's surface should be non-reflective.
10. The instruments should carry US FDA or European CE with a four-digit notified number.
11. The brand name, along with the catalog number, should be attached to the instruments.
12. The material used in scissors should be an alloy of Titanium & / Stainless steel/ Tungsten carbide, resulting in superior corrosion resistance and precise cutting with durable cutting.
13. All instruments should have catalog numbers embossed on them.
14. The firm will have to demonstrate the instruments if required.
15. All instruments quoted should be from the same manufacturer with international quality standards with a maximum 5 % variation in sizes may be allowed.
16. All surgical Instruments should be from the same country of origin and from the same manufacturer.
17. All instruments and equipment must be of high-grade medical quality stainless steel, reflection-free (matte finish), lightweight, and user-friendly (both for application & comfortable use) unless otherwise specified, in anti-glare style finishing, if the used material and the construction of the instrument will allow it. The matting shouldn't be affected in an electro-plating process.
18. Surgical Instruments should have the following certifications:
US FDA or European CE with four digits notified body.
DIN Certificate
ISO 9001: 2015
EN ISO 13485: 2016
EN ISO 7153
19. Instruments should be supplied with 15 large-size Sterilisation containers for storage as well as for sterilization purposes; the same container should be autoclavable along with the instruments.
20. The sterilization containers should meet international standards and be approved for steam sterilization procedures to EN 285: 2008 and validated according to ISO17665.
21. The sterilization container lid is made of high-performance plastic- PPSU.
22. The sterilization container should be resistant to permanent deformation.
23. The sterilization container should be lightweight and scratch-proof.
24. The sterilization container bottom is made of high-quality aluminum.
25. The sterilization container should have carrying handles with a 90-degree arrester made of fiberglass reinforced PEEK that prevents the jamming of the hands with a 90-degree arrester.
26. The sterilization container handle should have low heat conductivity and heat storage properties (burning of hands)
27. The sterilization container should have a frontal accommodation field for labels and color tags.
28. The lid of the sterilization container should have a microbial barrier filter to resist the passage of microbes.

Preferably
Ams
26/06/24

Am

my

maledupta
26/06/24

Page no. - 01 of Page

29. The microbial barrier filter should be reusable and easily accessible during manual and mechanical cleaning.
30. The sterilization container lid should have a silicone lip gasket with a functional safety design, which provides a hygienically perfect smooth surface and no cavities for particles/ germs.
31. The locking system of the sterilization container should have a hinge-less opening button.
32. The sterilization container should have a red/ green indicator- during the sterilization cycle, the indicator turns to green, and as soon as the opening button is pushed, the indicator turns to red.
33. The surgical instruments and sterilization containers should be from the same manufacturer (no third-party product will be accepted)
34. The sterilization containers must have the following certifications:
 DIN 58953-9
 ISO 17665-1
 ISO 11607
 EN 868
 US FDA or European CE with a four-digit notified body.
35. Wender should have experience in supplying similar surgical instruments to at least 3 other Institutes of National Importance

~~36. Warranty - 3 years (one time) - 2 years.~~ (m)

~~PM~~

36 Warranty of one year

PM
26/06/2024

Ph

meeta
26/06/2024