



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

F-No.-AIIMS/JMU/BOO/2024/LP/ 01

Dated: 7 September, 2024

Inviting Quotations for Purchase of Micropipettes for Microbiology Department, AIIMS, Vijaypur Jammu.

### QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of **Micropipettes** for **Microbiology Department** at AIIMS, Vijaypur, Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, **Vijaypur, District Samba, Jammu (184120)** up to **17-09-2024** till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of Equipments are given as under:-

S. No.	Name of the Equipment	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	Micropipettes	01					

#### **Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. Supply should be made within 15 days from the date of purchase order.
6. Price should be for Destination basis (i.e., concerned department).
7. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
8. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
9. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
10. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
11. Sealed quotation should be submitted by speed post/ copy courier/ by hand to Procurement Section, **AIIMS, Vijaypur, District Samba, Jammu (184120)** up to **17-09-2024** till 03:00 pm.
12. Vendors would get their sample check by user department through procurement section on last date of quotation opening.
13. Validity of quotation should be 90 days from the date of opening.
14. Sample to be submitted as and when required by the institute.
15. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

16. All column given in quotation should be filled otherwise quotation will not be accepted.
17. Along with quotation please mentioned Email-id and Contact Number.

*Raksha Kundal* 7 sep 24

**Dr Raksha Kundal**  
**Chairperson,**  
**Store Purchase Committee,**  
**AIIMS, Vijaypur, Jammu**

### **Specifications for Required Micropipettes:**

- Variable volume single channel Set of 6 (0.1-2.5 ul, 2-20 ul, 20-200 ul, 100-1000 ul, 10-100 ul and 0.5-10 ul) with Autoclavable tip Box & 96 tips
- All Pipettes should be supplied with a Carousel stand.
- Light weight, Mechanical pipette with Four-digit display.
- Pipette should come with two adjustment options – Calibration adjustment & Temporary adjustment.
- Quick connection clip to remove the lower part easily.
- Pipette Piston with highly resistant to heat, acids and alkalis, mildew, bleaches, aging, sunlight, and abrasion.
- Tip cone should have spring to reduce stress for up to 1 ml pipette.
- Volume adjustment with few turnarounds from maximum to min volume.
- Volume change should be available separately.
- Full pipette should be autoclaveable
- No discoloration upon UV irradiation.
- Viable calibration seal to indicate factory calibration not changed.
- Pipette specifications should meet the ISO 8655-2:2022 requirements.
- Pipette performance certificate should be provided along with the instrument.