



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

F-No.-AIIMS/JMU/BOO/2024/LP/ 02

Dated: | November, 2024

Inviting Quotations for Purchase of Air Oxygen blender required for the Department of Pediatrics, at AIIMS, Vijaypur Jammu.

### QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of Air Oxygen blender required for the Department of Pediatrics, Vijaypur Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, **Vijaypur, District Samba, Jammu (184120)** up to **8 -11-2024** till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of instruments are given as under:-

S. No.	Name of Items	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	Air Oxygen blender	01					

### **Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. Supply should be made within 15 days from the date of purchase order.
6. Price should be for Destination basis (i.e., concerned department).
7. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
8. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
9. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
10. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
11. Sealed quotation should be submitted by speed post/ copy courier/ by hand to Procurement Section, **AIIMS, Vijaypur, District Samba, Jammu (184120)** up to **8 -11-2024 till 03:00 pm.**
12. **Vendors would get their sample check by user department through procurement section before the last date of quotation opening.**
13. **Validity of quotation should be 90 days from the date of opening.**
14. **Sample to be submitted as and when required by the institute.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

15. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
16. All column given in quotation should be filled otherwise quotation will not be accepted.
17. Demo of the Item is must for quality check before the last date.
18. Please mentioned Email-id and Contact Number in the quotation.

Dr Raksha Kundal  
Chairperson,  
Store Purchase Committee,  
AIIMS, Vijaypur, Jammu

## Technical specifications

Name of the equipment	Technical specifications
<p>1. Air oxygen blender</p>	<ul style="list-style-type: none"> <li>• Should provide the mixed gas of up to 30 LPM.</li> <li>• Should combine compressed Medical Air and Oxygen (O<sub>2</sub>) to deliver blended pressurized gas at a precise oxygen concentration (FiO<sub>2</sub>) when set by the user.</li> <li>• Highly accurate delivery of FiO<sub>2</sub> with knob adjustment from 21 to 100%</li> <li>• Ideal for equipment requiring 2-30 LPM.</li> <li>• Non-interchangeable input port for air and oxygen.</li> <li>• Primary and auxiliary flow meter provisions provided.</li> <li>• ON/OFF functions ensures more accuracy during low output flow.</li> <li>• Highly audible pneumatic alarm when input pressure differentiates between 20±2 psi.</li> <li>• Rigid aluminium body and highly precise mechanical components ensures the accurate mixture of input gases.</li> <li>• Easy calibration makes it simple for end users.</li> <li>• Blender is suitable for respiratory applications including routine therapy, ventilator gas supply, Bubble CPAP, BiPAP, High flow oxygen therapy, Resuscitator and critically-limited NICU procedures.</li> <li>• Optional features of Air O<sub>2</sub> blender/mixer include: Input gas pressure regulators, Input gas filters and customized fittings for flow meter for Air O<sub>2</sub> blender and Air O<sub>2</sub> blender with flow meter (cascaded).</li> <li>• Maintenance: Warranty (2 years)</li> <li>• Should have the ISO certification for quality and safety</li> <li>• The unit should have BIS or US FDA or European CE certificate for the quoted model and certificate to be submitted.</li> <li>• Mandatory one visit each quarter during warranty</li> <li>• Training of hospital engineers &amp; staff, and demonstration is compulsory.</li> </ul>