



F-No.-AIIMS/JMU/BOO/2024/LP/03

Dated: | November, 2024

Inviting Quotations for Purchase of T piece resuscitator with blender required for the Department of Pediatrics at AIIMS, Vijaypur Jammu.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of **T piece resuscitator with blender** required for the Department of **Pediatrics**, Vijaypur Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, **Vijaypur, District Samba, Jammu (184120)** up to **8-11-2024** till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of instruments are given as under:-

S. No.	Name of Items	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	T piece resuscitator with blender	01					

Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. Supply should be made within 15 days from the date of purchase order.
6. Price should be for Destination basis (i.e., concerned department).
7. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
8. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
9. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
10. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
11. Sealed quotation should be submitted by speed post/ copy courier/ by hand to Procurement Section, **AIIMS, Vijaypur, District Samba, Jammu (184120)** up to **8-11-2024** till 03:00 pm.
12. **Vendors would get their sample check by user department through procurement section before the last date of quotation opening.**
13. **Validity of quotation should be 90 days from the date of opening.**
14. **Sample to be submitted as and when required by the institute.**
15. **Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

16. All column given in quotation should be filled otherwise quotation will not be accepted.
17. Demo of the Item is must for quality check before the last date.
18. Please mentioned Email-id and Contact Number in the quotation.

Raksha Kundal
1/11/24
Dr Raksha Kundal
Chairperson,
Store Purchase Committee,
AIIMS, Vijaypur, Jammu

Technical specifications

Name of the equipment	Technical specifications
<p>1. T piece resuscitator with blender</p>	<ol style="list-style-type: none"> 1. The device should be ideal for use in Labour room, NICU and during transportation. 2. Provide controlled and accurate PEEP and PIP for resuscitation and ventilation of newborns and infants upto 10 kg 3. It should be powered by gas flow with no electrical or battery operation. 4. Blender <ul style="list-style-type: none"> ○ Built-in active mechanical Air oxygen blender and able to deliver FiO₂ could be varied from 21% to 100% ○ Delivered FiO₂: Highly calibrated blender ensures accurate % of oxygen delivered with minimal variability. 5. It should be light weight and easy to handle. 6. In-built Pressure Gauge (manometer) to set & indicate delivery of PIP & PEEP <ul style="list-style-type: none"> ○ Manometer range-10 to 70 cm H₂O (mbar) with accurate accuracy 7. Peak Inspiratory Pressure (PIP) at 8 lpm: 4 to 36 cmH₂O 8. Positive End Expiratory Pressure (PEEP) at 8 lpm: 4 cmH₂O to 36 cmH₂O 9. Safety provision with adjustable Pressure Relief Value for maximum limiting 10. Maximum pressure relief 60 cm H₂O 11. Delivered oxygen up to 100% depending on gas supply. 12. Unit should be compatible with neonatal mask & endotracheal tube 13. The patient T-Piece should have port for surfactant delivery 14. Should be compatible for use with heated humidifier 15. Spiral Heated wire circuit with integrated T-Piece compatible for use with heated humidifier should be supplied with system. 16. Aluminium Metal casing for durability 17. Metal gas inlet and patient outlet connectors for durability 18. Gas inlet should fit standard oxygen line without any special tubing 19. Option of mounting on horizontal or vertical mounting bracket, IV poles, DIN or Dovetail Rail systems 20. Operating and storage limits - 20°C to 50°C and Up to 90 % relative humidity 21. ISO standards for medical devices: ISO 13485:2016 22. Supplies with each unit of Infant T-piece Resuscitator Unit <ul style="list-style-type: none"> ○ Supply line- Two sets with each equipment ○ Test Lung-One ○ T Piece Resuscitation Circuits Resuable 2 and disposable 10 pieces

	<ul style="list-style-type: none">○ Resuscitation Masks of size (medium) 0 size-10 Piece○ Resuscitation Masks of size (small) 00 size-10 Piece○ Resuscitation Masks of size 1 size-10 Piece <p>23. Demonstration of product prior to purchase is must.</p> <p>24. Original literature, and not thew photocopy, to be supplied with the quotation</p> <p>25. Warranty: For 2 years</p>
--	---

26. Should have *AIS / US - FAA / CE-CE certifiact*