



F-No.-AIIMS/JMU/BOO/2024-25/LP/ 22

Dated: 21 January, 2025

Inviting Quotations for Purchase of Equipment required for the Department of Pediatric Surgery, at AIIMS, Vijaypur Jammu.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of Equipment required for the Department of Pediatric Surgery, at AIIMS, Vijaypur Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, Vijaypur, District Samba, Jammu (184120) up to 21-01-2025 till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of Equipment are given as under:-

S. No.	Name of Equipment	Make/Model	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	Head Light		01					
2	Surgical Loupe		01					


Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. The aforementioned items must confirm to the specifications attached as Annexure-1.
6. Supply should be made within 15 days from the date of purchase order.
7. Price should be for Destination basis (i.e., concerned department).
8. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
9. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
10. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
11. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
12. Sealed quotation should be submitted by speed post/ copy courier/ by hand to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to 21-01-2025 till 03:00 pm.
13. **Vendors would get their sample check by user department through procurement section before the last date of quotation opening.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

14. Validity of quotation should be 90 days from the date of opening.
15. Sample to be submitted as and when required by the institute.
16. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
17. All column given in quotation should be filled otherwise quotation will not be accepted.
18. Demo of the Item is must for quality check before the last date.
19. Please mentioned Email-id and Contact Number in the quotation.


21/1/25
Dr Raksha Kundal
Chairperson,
Store Purchase Committee,
AIIMS, Vijaypur, Jammu

Pediatric surgery

“MAGNIFYING PRISMATIC LOUPE HEAD MOUNTED WITH HEADLIGHT”

MAGNIFYING PRISMATIC LOUPE Specifications:

- a) Magnification approx: 2.5 X Panoramic Prismatic
- b) Interpupillary distance adjustable or customized as per your requirement; Loupe customized as per near-sightedness also
- c) Configurations: TTL
- d) Field of view: Width and depth should be more than 7 TO 9 CM (Right angled view)
- e) Weight: 108 to 112gm
- f) Working distance: Customized as per Surgeon requirement.
- g) Waterproof, hypo allergenic and corrosion resistant
- h) Frame should be lightweight, adjustable nose pad and of international standards
- i) Supplied with accessories like **side splash protectors, cloth for cleaning, screw with keychain, box for loupe.**
- j) Should be manufactured by a well-known international firm with BIS/ISO/European CE (four digit notified body number) or US FDA certified
- k) Cost of customization and other accessories should be included.
- l) 2 year warranty for Loupe and from the date of invoice for any change in IPD and working distance.
← (plus 6 years CMC to be provided).
- m) Demonstration of the Item may be asked at any stage of procurement.

Head Light	
	1. Portable Battery operated Light source - Powerful lightweight LED light source with uniform field of distribution which should be easily mountable on the loupe.
	2. Total fitment over the loupes should be lightweight (Less than 20 grams)
	3. Spot Size should be ~ 70 mm at 300 mm (or better).
	4. Mountable Adapter for loupes - To attach to the surgical loupe (mentioned above)
	5. Soft case - Shock-proof protection for loupes, light source and accessories
	6. Light intensity - > 50000 Lux (at distance of 300 mm) or better
	7. Typical color temperature - Resembles daylight (5000-6000K)
	8. Batteries Charger - Medical grade lithium- ion smart battery packs for fast charging. It should be Processor controlled with country specific power adaptor. Battery run time - 8 hours or more at 100 % light intensity and 8 hours or more at 50% light intensity.
	9. Carrying Case to be provided.

TECHNICAL SPECIFICATION OF HEAD LIGHT

S.No	Item	Technical Specification
01.	LED Headlight with Cold Illumination	<ol style="list-style-type: none"> 1. Should be comfortable and mobile LED Head Light and easy touse without any cables 2. Should have a maximum luminous Intensity of around 2,00,000lux or more 3. Should have color temperature of 5000K 4. Should have minimum battery service life of 6 hours after fullcharge 5. Batteries should be capable of being charged easily withoutpowering off 6. Should be able to have rapid USG charging from wall adaptor orpersonal computer 7. Should have integrated battery status indicators 8. Should have continous digital dimming facility 9. Should be light weight and less than 600 gms 10. Should be able to focus down to about 65 mm 11. Should be comfortable through a removable, washable headband 12. Should have an extra spare battery 13. Should be provided with easily attachable 2 Dioptre and 3dioptre Magnifying Lenses 14. Shoud be European CE or US FDA approved and of International safety standards 15. Prior Demo is required

16. Warranty 1 year. or as per given by OEM.