



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

F-No.-AIIMS/JMU/BOO/2024/LP/16

Dated: 6 January, 2025

Inviting Quotations for Purchase of Equipments required for the Department of Blood Bank, at AIIMS, Vijaypur Jammu.

### QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of Equipments required for the Department of Blood Bank, at AIIMS, Vijaypur Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, Vijaypur, District Samba, Jammu (184120) up to 24-01-2025 till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of Equipments are given as under:-

S. No.	Name of Equipments	Make/Model	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	Tube Stripper		01					
2	Mobile Transport Box		01					

### **Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. The aforementioned items must confirm to the specifications attached as Annexure-1.
6. Supply should be made within 15 days from the date of purchase order.
7. Price should be for Destination basis (i.e., concerned department).
8. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
9. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
10. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
11. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
12. Sealed quotation should be submitted by speed post/ copy courier/ by hand to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to 24-01-2025 till 03:00 pm.
13. **Vendors would get their sample check by user department through procurement section before the last date of quotation opening.**
14. **Validity of quotation should be 90 days from the date of opening.**
15. **Sample to be submitted as and when required by the institute.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

16. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
17. All column given in quotation should be filled otherwise quotation will not be accepted.
18. Demo of the Item is must for quality check before the last date.
19. Please mentioned Email-id and Contact Number in the quotation.

*Raksha Kundal*  
*6/Jan/27*

**Dr Raksha Kundal**  
**Chairperson,**  
**Store Purchase Committee,**  
**AIIMS, Vijaypur, Jammu**

## Tube Stripper

1. It should have a stainless steel body with anti-rust properties.
2. Weight should not be more than 220 gms.
3. Should ensure the uniform pressure while pressing to close and automatic recoiling of the spring to release the handle for opening.
4. Should have Screwless rollers to avoid loosening of the rollers.
5. The roller must be of Derlin.
6. Should be capable of stripping, sealing/crimping, and have extra sharp cutting edges.
7. Should be an ergonomically designed handle for better grip.
8. It Should have a roller guide to avoid any damage to the tube.
9. There Should be provision for manual tube sealing by aluminum rings.
10. Original literature on equipment should be submitted.
11. Can be sterilized/autoclaved.
12. Should be ISO 9001:2015, ISO 13485-approved product.
13. The equipment should have US FDA or European CE or BIS Certifications.
14. The warranty period should be ~~two~~ *one year* *only per OEM/Reseller* years.
15. Necessary spare parts and accessories to be provided with the eqi ,
16. Should provide a user list with a satisfactory report from three reputed government institutes of national importance with contact details.
17. The company should have a market standing of more than 10 years.
18. Should provide a user list with a satisfactory report from three reputed government institutes of national importance.

### **Mobile Transport Box (Refrigerated Blood Transport Box)**

1. Meant specifically for transport of blood components including FFP
2. Both manufacturer and distributor/service provider should be ISO 13485 compliant.
3. Should be compliant with European CE or US FDA or BIS for this specific purpose.
4. Electrical safety conforms to standards for electrical safety IEC-60601
5. Lid should be fully insulated and fitted up with a perimetric rubber gasket, with a special locking device (granting a perfect seal).
6. Internal partitioning and securing should be possible for easy handling and preventing damage to fragile FFP units during tilting/harsh transport conditions.
7. Temperature range: infinitely adjustable between +10 C to -18C
8. Adjustable thermostat should be present to set for different temperatures for different transport functions eg +4°C for RBC and -18°C for FFP, and the present temperature and set temperature both should be displayed.
9. Cooling unit should have a hermetically sealed compressor and should be industrial grade granting the maximum reliability and safety during transport.
10. Should be able to store at least 150 bags.
11. CFC free insulation
12. Features should facilitate easy cleaning like round internal edges
13. For easy handling of the portable refrigerator there should be handles and there should either be inbuilt wheels or an attachable trolley. Must be compatible with the hospital transport vehicles in terms of electrical connections(Mobile transport van, Blood mobile etc)
14. Features for properly securing blood components without damaging them inside the container.
15. Facility to validate cold chain by Recording set transport temperature
16. Five years onsite warranty inclusive of all Spares and Labour etc. and Five years CMC to be quoted.
17. Voltages: both 12/24 V and 220-230V/1 phase /50 Hz
18. Connecting cables (included): for both the voltage (12/24V and 220-230V).
19. The vendor should provide IQ, OQ & PQ free of cost and PQ should be finalized after 3 months of observation.
20. Calibration certificate should be provided by the vendor during the warranty & CMC period without any extra cost.
21. Validation and calibration reports should have traceability to applicable national and international standards.
22. Complete with comprehensive set of spare parts and a suitable capacity voltage stabilizer with the charging set.
23. The make, rating, model, description, specifications, price quantity of each item should be furnished separately.
24. Necessary catalogues, technical write up in English, should be attached with the offer both in hard and electronic copies.
25. Performance, efficiency, other factors as applicable should be furnished.
26. Should provide electronic and hard copies of User Manual (English), Service manual (English) and Complete construction details with respect to material specification, thickness, finish etc.
27. Should provide a set of equipments for calibration (eg thermometer) and routine Preventive Maintenance as per manufacturer documentation in service/technical manual.
28. Should provide Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.
29. When there is a breakdown, the firm will carry out the repair within 24-48 hours
30. The company should have a market standing of more than 10 years.
31. Should provide a user list with a satisfactory report from three reputed government institutes of national importance with contact details.