



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

F-No.-AIIMS/JMU/BOO/2024-25/LP/18

Dated: 15 January, 2025

Inviting Quotations for Purchase of Equipment/Item required for the Department of Pediatrics, at AIIMS, Vijaypur Jammu.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of Equipment/item required for the Department of Pediatrics, at AIIMS, Vijaypur Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, Vijaypur, District Samba, Jammu (184120) up to 22-01-2025 till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of Equipment/Item are given as under:-

S. No.	Name of Equipment/Item	Make/Model	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	Neonatal vein finder (Diaphanoscope)		01					


Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. The aforementioned items must confirm to the specifications attached as Annexure-1.
6. Supply should be made within 15 days from the date of purchase order.
7. Price should be for Destination basis (i.e., concerned department).
8. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
9. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
10. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
11. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
12. Sealed quotation should be submitted by speed post/ copy courier/ by hand to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to 22-01-2025 till 03:00 pm.
13. **Vendors would get their sample check by user department through procurement section before the last date of quotation opening.**
14. **Validity of quotation should be 90 days from the date of opening.**
15. **Sample to be submitted as and when required by the institute.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

16. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
17. All column given in quotation should be filled otherwise quotation will not be accepted.
18. Demo of the Item is must for quality check before the last date.
19. Please mentioned Email-id and Contact Number in the quotation.


Dr Raksha Kundal
Chairperson,
Store Purchase Committee,
AIIMS, Vijaypur, Jammu

Technical specifications

Name of the equipment	Technical specifications
1. Neonatal vein finder (Diaphanoscope)	<ol style="list-style-type: none"> 1) Probe based device with embedded red and orange LEDs. 2) Should allow easy visualization of veins as well as arteries. 3) Should be suitable for use in preterm and term neonates. 4) Should have orange and red high-performance LED each equipped with a dimmer. 5) Can be used for diagnosis of pneumothorax or hydroceles. 6) Energy source: Battery operated 7) Accessories: 8) Main Unit 9) Charger 10) Carry Case 11) Operation Manual 12) Disinfection using standard procedures 13) Warranty: 1 year 14) Should be BIS or European CE or US FDA approved product and the certificate must be submitted 15) Manufacturer/ Supplier should have ISO certification for quality standards 16) Physical demonstration for technical evaluation of equipment is mandatory

