



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

Dated: 24.01.2025

**Corrigendum-3**

**Subject: Corrigendum for procurement of Vitals Equipments required for the Department of Endocrinology & Metabolism, AIIMS, Vijaypur, Jammu.**

**Ref.:** 1. CFQ No. AIIMS/JMU/BOO/2024/LP/07 dated 03<sup>rd</sup> December, 2024.  
2. Corrigendum-1 dated 20.12.2024.  
3. Corrigendum-2 dated 03.01.2025.

In context to the subject and reference cited above, it is to inform that the date for receipt of quotations for aforementioned calling for quotation is hereby extended upto 10.02.2025. Rest all the terms & conditions shall remain same.

**Encl.: As referred above.**

*Raksha Kundal*  
24/1/25

**Dr. Raksha Kundal,  
Chairperson,  
Store Purchase Committee,  
AIIMS, Vijaypur, Jammu**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)  
\*\*\*\*\*

**Dated: 03.01.2025**


**Corrigendum-2**

**Subject: Corrigendum for procurement of Vitals Equipments required for the Department of Endocrinology & Metabolism, AIIMS, Vijaypur, Jammu.**

**Ref.:** 1. CFQ No. AIIMS/JMU/BOO/2024/LP/07 dated 03<sup>rd</sup> December, 2024.  
2. Corrigendum-1 dated 20.12.2024.

In context to the subject and reference cited above, it is to inform that the date for receipt of quotations for aforementioned calling for quotation is hereby extended upto 21.01.2025. Rest all the terms & conditions shall remain same.

**Encl.: As referred above.**

  
**Dr. Raksha Kundal,  
Chairperson,  
Store Purchase Committee,  
AIIMS, Vijaypur, Jammu**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

**Dated: 20.12.2024**

**Corrigendum-1**

With reference to the Calling for Quotation(CFQ) F. No. AIIMS /JMU/BOO/2024/LP/07 dated 03<sup>rd</sup> December, 2024, regarding procurement of Equipments for Endocrinology & Metabolism Department, at AIIMS, Vijaypur Jammu, the last date of receiving of quotations was 20.12.2024, it is to inform that the date for receipt of quotations for aforementioned CFQ is hereby extended upto 2 .01.2025. Rest all the conditions are same.

**Dr. Raksha Kundal,  
Chairperson,  
Store Purchase Committee,  
AIIMS, Vijaypur, Jammu**

**Encl.: Previous quote invite.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

F.No.-AIIMS/JMU/BOO/2024/LP/ 07

Dated: 3 December, 2024

Inviting Quotations for Purchase of Vitals Equipments required for the Department of Endocrinology & Metabolism, at AIIMS, Vijaypur Jammu.

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for Purchase of Vitals Equipments required for the Department of Endocrinology & Metabolism, at Vijaypur Jammu. The quotation with copy of certificate of GST & other documents should be submitted to Procurement Section at AIIMS, Vijaypur, District Samba, Jammu (184120) up to 20-12-2024 till 03:00 pm. The quotations will be opened on the same day at 04:00 pm. Details of instruments are given as under:-

S. No.	Name of Equipments & Consumables	Make/Model	Qty.	Packs Size/Unit	HSN Code	Basic Price	GST %	Total Cost Inclusive of GST
1	Cardiac Autonomic Neuropathy Analyser		01					
2	Neuropathy Analyser (Vibrotherm)		01					

**Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. GST, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e., GST number and relevant document should be submitted along with quotation.
4. GST/other taxes please may be included in your quotation with the price in separate column.
5. Supply should be made within 15 days from the date of purchase order.
6. Price should be for Destination basis (i.e., concerned department).
7. Payment will be released after certification from HOD of Concerned Department / Inspection committee of AIIMS, Vijaypur, Jammu.
8. **Quotation Name, Name of Department and No. must be mentioned on top of envelope.**
9. Liquidated damage shall be @ 0.5% for delayed supply per week or part of week for delay subject to maximum of 10%.
10. AIIMS, Vijaypur, Jammu reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Jammu reserves the right to increase/ decrease the number of required quantities all other terms & condition.
11. Sealed quotation should be submitted by speed post/ copy courier/ by hand to Procurement Section, AIIMS, Vijaypur, District Samba, Jammu (184120) up to 20-12-2024 till 03:00 pm.
12. **Vendors would get their sample check by user department through procurement section before the last date of quotation opening.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, VIJAYPUR, JAMMU-184120  
(A Central Autonomous Body under PMSSY, MoH&FW, Government of India)

\*\*\*\*\*

13. Validity of quotation should be 90 days from the date of opening.
14. Sample to be submitted as and when required by the institute.
15. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
16. All column given in quotation should be filled otherwise quotation will not be accepted.
17. Demo of the Item is must for quality check before the last date.
18. Please mentioned Email-id and Contact Number in the quotation.

**Dr Raksha Kundal**  
Chairperson,  
Store Purchase Committee,  
AIIMS, Vijaypur, Jammu

## Cardiac Autonomic Neuropathy System ( CANS ) specifications

- Should analyse Sympathetic and Parasympathetic autonomic nervous function
- Simple six sequence tests to complete the comprehensive study
- Should have fully automatic NIBP (Non-Invasive Blood Pressure) kit for reliable NIBP measurements
- Display and print of raw ECG and Heart Rate Histogram to support the diagnosis
- Fully automated WINDOWS PC based software
- Should have provision for Gain adjustment 1x or 2x
- Software should support all versions of windows operated computer
- Two full A4 page colour report with patient details
- Resting ECG with resting HR
- Deep Breathing ECG, Expiration Inspiration ratio recording
- Supine to Standing ECG, R-R interval and 30:15 ratio recording
- Valsalva Manoeuvre ECG, R-R interval, Valsalva ratio recording and graph should also display start and end point of Valsalva measurement in the report.
- Postural Hypertension resting BP, BP immediately after standing and 1or 2minute after standing, with Fall in systolic BP assessment
- Sustained Hand grip BP before grip and BP after 5 minutes of sustained grip. Increase in diastolic BP assessment.
- Report print out with final Impression consisting of Parasympathetic function, Sympathetic function and Cardiac Autonomic Neuropathy function.
- Should be compatible to work with Windows 7,8 & 10 based computer
- Report should have 1-minute ECG waveform with R wave marker for all the postures
- Protocols based on International Scientific papers by D.J. Ewing et all and Aaron I. Vinik.
- Should be an ISO13485:2016 certified company

### List of accessories:

1. 3 lead ECG patient cable – 1 No.
2. Adult normal BP cuff – 1 No.
3. Valsalva Maneuver Holder with tube – 1 No.
4. Hand Grip unit with wire – 1 No.
5. PVC mouth pieces for Valsalva – 25 Nos.
6. Disposable ECG Electrodes – 50 Nos
7. Reusable ECG electrodes – 5 Nos
8. Mains Cable 2-meter length – 1 No.
9. USB Interface Cable – 1 No.
10. Earth Cable – 1 No.
11. User Manual
12. Software CD
13. Fuse 400mA, S/B – 4 Nos.

- Warranty period of three years (3 years)
- Should be ISO Certified.

### Specifications of Neuropathy analyzer for Vibration, Hot and cold sensory perception

- Should have the Vibration Frequency of 50 Hz
- The Vibration output should vary between 0 and 50 Volts
- Should have LED display read out
- Vibration Probe tip should have 15 mm diameter and made of ABS
- Temperature Range should be 1 – 50 °C in steps of 1 °C and rate of change should be 1 °C per second
- There should be an alarm for Lower Cut-off at 1 °C and High cut-off 50 °C
- Probe should have thermoelectric module with heat sink
- Thermal Probe tip contact should be 20 mm diameter
- A wired remote key pad should be provided
- The main unit should also have a key pad
- Should have USB computer connectivity
- Software supplied should support all versions of Windows
- Patient data should be stored in the computer
- Power consumption should be 150VA
- Weight should be less than 10Kgs
- warranty period up to three (3) years
- should be ISO Certified